




Maestro

SOFTWARE

DESCRIPTION

Maestro is a software for managing patient data that can interface with all the Inventis device, controlling or getting exams results from them.

OPERATING MODES

- Noah module
- OtoAccess module
- Stand-Alone with proprietary database
- Stand-Alone without database functions (file based)
- Third-party Applications (exchange data with external applications): MII proprietary protocol and GDT standard

AVAILABLE TESTS

- Audiology exams
- REM
- HIT
- Evoked Potentials
- Video-Otoscopy
- Video and Images acquisition
- Questionnaire
- Session Data

COMPATIBLE DEVICES

- Triangle (Audiometer)
- Bell (Audiometer)
- Harp (Audiometer)
- Piano (Audiometer)
- Piccolo (Audiometer)
- Cello (Audiometer)
- Trumpet (REM, Audiometer)
- Drum (HIT)
- Flute (Middle-Ear Analyzer)
- Clarinet (Middle-Ear Analyzer)
- Viola (Middle-Ear Analyzer, Audiometer)
- Timpani (Handheld Tympanometer, Audiometer)
- Celesta (Evoked Potentials system)
- Delfino (Otoscope)
- Harmonica (Otoscope)
- MediCam (Video-endoscopy camera)
- MediCam HD (HD Video-endoscopy camera)
- Flexia 2 (Video-Nasopharyngoscope)
- Analog-to-USB video converter
- HDMI-USB HD video converter

COMPATIBLE ACCESSORIES

- Maestro Control Panel

INTERFACE WITH EQUIPMENT

USB connection

PC MINIMUM REQUIREMENTS

CPU: Intel® i5, 6th generation or above, 8 GB RAM

Operating System: Windows 10 and Windows 11, 64 bits

Display resolution: suggested 1920 x 1080 or higher (zoom factor 100% or 125%)

Local network (only for db installation in network folder): 1 Gb/s

Free space on HD: 3 GB

NOAH MINIMUM REQUIREMENTS

NOAH 4.4 or greater (or NOAH Engines supporting NOAH 4.4 or greater)

OTOACCESS® MINIMUM REQUIREMENTS

OtoAccess2 version 2.4.1.18 or greater

SOFTWARE SUPPORT

Remote assistance tool: included

SOFTWARE UPDATES

Available for free

MII INTERFACE

Library for interfacing Maestro with 3rd party software

Available upon request (IT technician intervention requested)

INCLUDED PARTS

- Inventis Software Suite
- Electronic IFU

APPLICABLE STANDARDS

Software life cycle: IEC 62304

Health software: IEC 82304-1, IEC 81001-5-1

CE CERTIFICATE

MDR 2017/745/EU Classification: Class IIa

Classification rule (Annex VIII, 2017/745): 11

Notified body: TÜV SÜD Product Service GmbH [0123]

Maestro is manufactured by:

INVENTIS S.r.l.

VIA AUSTRIA 38

35127 PADOVA – ITALY

PHONE: +39.049.8962 844

info@inventis.it

www.inventis.it

Follow us on LinkedIn

<https://www.linkedin.com/company/inventis-srl>

The Inventis Quality System complies with ISO 13485 standard.

Inventis® is a registered trademark of INVENTIS S.r.l.